FirstTemp* Genius*

Infrared Thermometer

User Guide

Follow these simple instructions to ensure maximum performance from your GENIUS thermometer

- Remove the **GENIUS** thermometer from its base unit by lifting firmly near the RELEASE/EJECT button.
- Note the equivalence setting on display.
- Ensure tympanic mode is selected by pressing MODE button.

- Place probe tip in ear canal and seal opening.
- In infants, seal outer opening of canal.
- Aim **GENIUS** thermometer towards the tympanic membrane.

- Press blue RELEASE/EJECT button to discard the used probe cover.

- Check the probe lens is clean.
- Remove any specs with a dry cloth.
- **DO NOT USE** if probe tip is dirty.

- Press and release the SCAN button.
- Remove probe tip from ear as soon as triple bleep is heard and display flashes DONE.

- Return the thermometer to base unit and clip in place for storage.

- Place a **new** probe cover on the probe tip.
- Never re-use covers, as low readings could be recorded.
- Ensure the probe cover is tightly in position on the tip.

- Read the temperature displayed in the window.

ALWAYS REMEMBER

1. **Always use a new probe cover.**
2. **Keep probe lens clean.**
3. **Remove the GENIUS thermometer from its base with care and in the proper manner.**

Note: If the **GENIUS** thermometer has a probe cover installed but is left unused for more than approximately 30 seconds, it enters a power saving SLEEP mode. Installing a **new** probe cover will allow the next temperature to be taken.

* Trademark of Sherwood Services AG. © 2004 Tyco Healthcare Group LP. All rights reserved.

P-TH-P-FTGenius02/GB

tyco Healthcare

Tyco Healthcare UK Limited
154 Fareham Road
GOSPORT
HAMPSHIRE
UK - PO13 0AS
Tel: +44 (0) 13 29 22 41 59
Fax: +44 (0) 13 29 22 40 83
www.tycohealthcare.co.uk

03/2004